



# Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 1 / 75

## INTRODUCTION

---

This document provides a summary of the design verification activities and results for the AIRone emergency ventilator. This includes verification of design specifications, safety tests, and usability tests.

The design specifications are described in detail in the Design Specifications Document. For each specification, the applicable test reports are summarized in this document.

Usability tests and safety tests are also summarized in this document.

## TABLE OF CONTENTS

---

Introduction.....	1
Table of Contents.....	1
Conclusion .....	1
Usability Tests.....	2
03-04-2020 .....	2
10-04-2020 .....	2
Safety Tests.....	2
Electrical Safety .....	2
Electromagnetic Compatibility .....	2
Design Specifications .....	3
Test Reports.....	10

## CONCLUSION

---

The AIRone is currently undergoing the final stages of design verification. Almost all design specifications have been verified. Usability testing has yielded some improvements to the design, which have all been implemented and verified as well. Electrical safety has been verified. Although full compliance with the standards regarding Electromagnetic compatibility has not been acquired, we believe that the AIRone is ready for clinical testing when needed in emergency situations where no conventional ventilation equipment is available.



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 2 / 75

## USABILITY TESTS

---

Two formative usability test sessions were held with several intensive care physicians and ventilation specialists. Special focus was paid to ease of use, safety, and applicability for COVID-19 patients on intensive care units in The Netherlands.

### 03-04-2020

#### *Participants:*

██████████, Intensivist; ██████████, Intensivist; ██████████, Technical Physician; ██████████, Technical Physician

#### *Conclusion:*

The device is intuitive to use. The current alarms are sufficient, but the default alarm limits should be smaller. COVID-19 patients generally require complex ventilation strategies, so the device must be able to handle high breathing rates up to 40/min with large minute volumes. Additionally, the possibility to perform expiratory holds is a must. Also a way to detect AutoPEEP should be added.

### 10-04-2020

#### *Participants:*

██████████, Intensivist; ██████████, Technical Physician; ██████████, Ventilation Practitioner

#### *Conclusion:*

The user interface works as it should, comparable to standard ventilators. Starting ventilation is easy. The device is able to provide the ventilation performance that is required. Peak pressure is generally set as relative pressure above PEEP instead of relative to ambient pressure. Pressure settings and especially PEEP should be adjustable in small intervals of 1 or 2 cmH2O. The time-axis of the flow and pressure curves should be smaller, generally 4 to 5 breaths should be displayed.

## SAFETY TESTS

---

### ELECTRICAL SAFETY

Leakage current testing was performed using the acceptance criteria for leakage current as defined in IEC 60601-1:2006. Measured leakage current was within acceptable limits for all tests.

### ELECTROMAGNETIC COMPATIBILITY

Explorative tests were performed for radiated immunity, radiated emission, and electrostatic discharge. Tests and acceptance criteria were based on IEC 60601-1-2:2015.

Radiated immunity testing indicated that some components were sensitive for certain irradiated frequencies. The speaker emitted noise when irradiated at 54, 117 and 124 MHz. The UPS module started vibrating when

	<b>Design Verification</b>	
	OPERATIONAIR: OperationAIR (OperationAIR)	Date: 2020/04/29 14:28:42
	Document: 4.1	Page: 3 / 75

irradiated with frequencies around 270 MHz. And the screen malfunctioned when irradiated in the 200 - 235 MHz range. The following design changes were made to eliminate these sensitivities:

- The speaker cable was placed directly on the protective earthing and capacitively coupled using conductive material
- Ferrite filters were placed on the cables to and from the UPS module
- The cable to the screen was shielded using conductive material and this shielding was connected to the protective earthing
- The screen was directly connected to the protective earthing

Radiated emission between 30 MHz and 1 GHz was measured in an anechoic chamber. Quasi-Peak emission limits were exceeded in two frequency bands: 42.7 dB $\mu$ V/m around 40 MHz, and 51.9 dB $\mu$ V/m at 219 MHz. Although these results technically constitute a failure, we consider the device acceptable for emergency use. Continued development is required.

Testing for bursts of electrostatic discharges at 8 kV showed that the flowsensor was susceptible. Discharges to the screws on the front of the device, close to the location of the flowsensor, could disrupt the sensor data communication. For this reason, the screws close to the flowsensor that were accessible from the exterior were replaced with non-conductive screws.

## DESIGN SPECIFICATIONS

Item	Tested By	Execution Date	Last Modification	Test Result
<b>SPEC-31</b> Ventilation mode	<b>XTC-30</b> Error margins: Pressure loss (TC-76)	2020/04/02	2020/04/02 08:21:46	passed
	<b>XTC-49</b> Pressure Rise Time (TC-120)	2020/04/05	2020/04/05 13:14:49	failed
	<b>XTC-38</b> Peak-plateau difference (TC-121)	2020/04/02	2020/04/02 16:57:29	passed
<b>SPEC-46</b> Usability user-interface	<b>XTC-110</b> Usability test (TC-65)	2020/04/10	2020/04/24 09:47:51	documentation
<b>SPEC-34</b> Visualized measurements	<b>XTC-2</b> Visualized measurements (TC-31)	2020/03/27	2020/03/27 17:41:48	passed
<b>SPEC-150</b> Instructions For Use	<b>XTC-127</b> IFU packed with device (TC-162)	2020/04/24	2020/04/24 10:41:32	passed
<b>SPEC-42</b> Compatibility with standard hospital equipment	<b>XTC-77</b> Compatibility with standard hospital equipment (TC-39)	2020/04/06	2020/04/06 18:59:48	passed
<b>SPEC-141</b> Inspiratory hold	<b>XTC-65</b> Inspiratory Hold (TC-156)	2020/04/10	2020/04/10 07:45:23	passed with deviation



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 4 / 75

Item	Tested By	Execution Date	Last Modification	Test Result
<b>SPEC-156</b> Expiratory Hold	<b>XTC-111</b> Expiratory Hold (TC-170)	2020/04/24	2020/04/24 08:03:36	passed
<b>SPEC-73</b> Interruption of power supply	<b>XTC-82</b> Back up battery (TC-10)	2020/04/07	2020/04/10 08:03:11	passed
<b>SPEC-60</b> Breaks	<b>XTC-66</b> IFU: Use breaks when needed (TC-104)	2020/04/06	2020/04/06 10:01:15	passed
	<b>XTC-51</b> Breaks (TC-103)	2020/04/05	2020/04/06 17:17:52	passed
<b>SPEC-27</b> Flow Sensor	<b>XTC-56</b> Flow sensor (TC-24)	2020/04/05	2020/04/05 11:57:41	passed
<b>SPEC-153</b> HEPA filter	<b>XTC-112</b> HEPA filter (TC-171)	2020/04/24	2020/04/24 08:11:37	passed
<b>SPEC-25</b> HME-filter	<b>XTC-14</b> HME filter (TC-70)	2020/03/31	2020/03/31 17:15:06	passed
<b>SPEC-166</b> Ventilator	<b>XTC-123</b> Ventilator (TC-172)	2020/04/24	2020/04/24 10:00:19	passed
<b>SPEC-155</b> Ppeak setting	<b>XTC-109</b> Pressure setting (TC-168)	2020/04/23	2020/04/23 09:42:42	passed
<b>SPEC-35</b> PEEP setting	<b>XTC-43</b> PEEP value (TC-30)	2020/04/06	2020/04/06 13:40:19	passed
	<b>XTC-46</b> PEEP stability (TC-133)	2020/04/06	2020/04/06 14:09:20	passed
	<b>XTC-60</b> PEEP settings (TC-32)	2020/04/05	2020/04/05 13:18:51	passed
<b>SPEC-20</b> Breathing rate setting	<b>XTC-25</b> Error margins: Breathing rate (TC-78)	2020/04/06	2020/04/06 13:25:22	passed
<b>SPEC-38</b> I:E ratio setting	<b>XTC-27</b> Error Margins: I:E ratio (TC-80)	2020/04/02	2020/04/24 08:22:50	failed
	<b>XTC-69</b> I:E ratio (TC-148)	2020/04/24	2020/04/24 08:13:07	passed
<b>SPEC-37</b> FiO2 setting	<b>XTC-57</b> FiO2 setting (TC-34)	2020/04/05	2020/04/05 12:01:25	passed
<b>SPEC-6</b> Alarm for Inspiratory O2 concentration (FiO2)	<b>XTC-86</b> Oxygen alarm (TC-9)	2020/04/07	2020/04/07 09:09:38	passed
	<b>XTC-87</b> Air pressure drop alarm (TC-155)	2020/04/07	2020/04/07 09:19:03	passed

All dates and times are in the project's default time zone and formatting.



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 5 / 75

Item	Tested By	Execution Date	Last Modification	Test Result
<b>SPEC-9 Alarm for positive end-expiratory pressure (PEEP)</b>	<b>XTC-101 PEEP alarm (TC-8)</b>	2020/04/13	2020/04/13 09:12:01	passed
<b>SPEC-8 Alarm for Tidal volume (Tv)</b>	<b>XTC-124 Tidal volume alarm I (TC-7)</b>	2020/04/24	2020/04/24 10:18:02	passed
	<b>XTC-125 Tidal volume alarm II (TC-68)</b>	2020/04/24	2020/04/24 10:19:26	passed
<b>SPEC-89 Alarm for Peak Pressure (Ppeak)</b>	<b>XTC-102 Plateau pressure alarm II (TC-83)</b>	2020/04/13	2020/04/13 09:19:16	passed
<b>SPEC-87 Alarm for Apnea or disconnect</b>	<b>XTC-103 Apnea/Disconnect alarm (TC-101)</b>	2020/04/13	2020/04/13 09:33:34	passed
<b>SPEC-84 Alarm for empty battery</b>	<b>XTC-128 Back-up battery empty alarm (TC-84)</b>	2020/04/24	2020/04/24 10:57:08	failed
<b>SPEC-19 Clear explanation per alarm on GUI</b>	<b>XTC-113 Clear explanation per alarm (TC-17)</b>	2020/04/24	2020/04/24 08:59:03	passed
<b>SPEC-95 Error margins: Tidal Volume</b>	<b>XTC-28 Error margins: Tidal Volume (TC-81)</b>	2020/04/02	2020/04/06 20:13:57	passed with deviation
<b>SPEC-94 Error margins: I:E ratio</b>	<b>XTC-27 Error Margins: I:E ratio (TC-80)</b>	2020/04/02	2020/04/24 08:22:50	failed
<b>SPEC-93 Error margins: FiO2</b>	<b>XTC-26 Error margins: FiO2 (TC-79)</b>	2020/04/02	2020/04/02 09:32:59	passed
<b>SPEC-92 Error margins: breathing rate</b>	<b>XTC-25 Error margins: Breathing rate (TC-78)</b>	2020/04/06	2020/04/06 13:25:22	passed
<b>SPEC-90 Error margins: Lung/airway pressure</b>	<b>XTC-30 Error margins: Pressure loss (TC-76)</b>	2020/04/02	2020/04/02 08:21:46	passed
<b>SPEC-36 FiO2 range</b>	<b>XTC-26 Error margins: FiO2 (TC-79)</b>	2020/04/02	2020/04/02 09:32:59	passed
<b>SPEC-26 Tidal Volume</b>	<b>XTC-34 Tidal volume range (TC-23)</b>	2020/04/02	2020/04/02 09:19:27	passed
<b>SPEC-40 Inspiratory Flow</b>	<b>XTC-70 Inspiratory flow (TC-37)</b>	2020/04/06	2020/04/07 11:44:59	passed
<b>SPEC-118 Reliability</b>	<b>XTC-130 Measurement Reliability (TC-106)</b>	2020/04/24	2020/04/24 11:09:06	passed
<b>SPEC-128 14 day reliability</b>	<b>XTC-79 Long-term reliability (TC-38)</b>	2020/04/06	2020/04/08 11:01:57	passed
	<b>XTC-81 Specification of expected durability (TC-130)</b>	2020/04/07	2020/04/07 08:07:49	passed with deviation

## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 6 / 75

Item	Tested By	Execution Date	Last Modification	Test Result
<b>SPEC-29</b> Peak pressure	<b>XTC-106</b> Fail-safe valve (TC-67)	2020/04/10	2020/04/15 11:22:04	passed
<b>SPEC-146</b> Flow sensor shielding	<b>XTC-131</b> Flow Sensor (TC-169)	2020/04/24	2020/04/24 11:10:20	passed
<b>SPEC-144</b> Airflow materials	<b>XTC-114</b> Airflow materials (TC-159)	2020/04/24	2020/04/24 09:01:40	passed
<b>SPEC-129</b> Robustness	<b>XTC-91</b> Push test (TC-61)	2020/04/08	2020/04/08 09:38:51	passed
	<b>XTC-92</b> Impact test (TC-62)	2020/04/08	2020/04/08 09:39:39	passed
	<b>XTC-120</b> Internal components tightly attached (TC-4)	2020/04/24	2020/04/24 09:46:59	passed with deviation
	<b>XTC-68</b> Tipping test (TC-109)	2020/04/06	2020/04/06 11:15:24	passed
	<b>XTC-118</b> Tipping test II (TC-110)	2020/04/24	2020/04/24 09:38:40	passed with deviation
	<b>XTC-11</b> Do not push warning (TC-129)	2020/03/31	2020/03/31 15:58:06	passed
<b>SPEC-126</b> Cleaning and disinfection	<b>XTC-115</b> Cleaning and disinfection (TC-127)	2020/04/24	2020/04/24 11:12:51	passed
<b>SPEC-120</b> Alarms are easy to understand and intuitive	<b>XTC-97</b> Alarms easy to understand (TC-119)	2020/04/08	2020/04/08 11:56:22	passed
<b>SPEC-113</b> Impairment of cooling	<b>XTC-94</b> Impairment of cooling (TC-99)	2020/04/06	2020/04/08 11:18:41	passed
<b>SPEC-103</b> No air leakage	<b>XTC-116</b> Pressure Hold (TC-91)	2020/04/24	2020/04/24 09:20:20	failed
<b>SPEC-100</b> External exhaust outlets	<b>XTC-58</b> External exhaust outlet oxygen rich environment (TC-86)	2020/04/05	2020/04/05 12:13:52	passed
<b>SPEC-99</b> Temperature of inspired air	<b>XTC-93</b> Temperature of applied parts (TC-85)	2020/04/08	2020/04/08 11:05:25	passed
<b>SPEC-85</b> High Voltage circuit outside casing	<b>XTC-9</b> 230 V circuit outside casing (TC-112)	2020/03/30	2020/03/30 11:14:35	passed
<b>SPEC-83</b> Testing of the fail-safe valve during production	<b>XTC-134</b> Fail safe valve test during production (TC-111)	2020/04/28	2020/04/29 07:11:46	passed



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 7 / 75

Item	Tested By	Execution Date	Last Modification	Test Result
<b>SPEC-80</b> Housing impact test	<b>XTC-92</b> Impact test (TC-62)	2020/04/08	2020/04/08 09:39:39	passed
<b>SPEC-79</b> Housing mechanical resistance	<b>XTC-91</b> Push test (TC-61)	2020/04/08	2020/04/08 09:38:51	passed
<b>SPEC-72</b> Leakage current testing	<b>XTC-63</b> Leakage current testing (TC-54)	2020/04/10	2020/04/10 08:02:28	passed
<b>SPEC-71</b> Patient leakage current	<b>XTC-63</b> Leakage current testing (TC-54)	2020/04/10	2020/04/10 08:02:28	passed
<b>SPEC-70</b> Touch current	<b>XTC-63</b> Leakage current testing (TC-54)	2020/04/10	2020/04/10 08:02:28	passed
<b>SPEC-67</b> Acoustic Energy	<b>XTC-19</b> Sound test (TC-49)	2020/04/06	2020/04/06 20:18:18	passed
<b>SPEC-65</b> Sharp edges	<b>XTC-117</b> Sharp edges (TC-48)	2020/04/24	2020/04/24 09:36:21	passed
<b>SPEC-63</b> Instability unintended force	<b>XTC-118</b> Tipping test II (TC-110)	2020/04/24	2020/04/24 09:38:40	passed with deviation
<b>SPEC-62</b> Instability incline surface	<b>XTC-68</b> Tipping test (TC-109)	2020/04/06	2020/04/06 11:15:24	passed
<b>SPEC-143</b> Visibility User Interface	<b>XTC-119</b> Usability with protective equipment (TC-41)	2020/04/24	2020/04/24 09:40:03	passed
<b>SPEC-145</b> Watchdog program	not covered by selected test run(s)			
<b>SPEC-53</b> Power Supply	<b>XTC-74</b> Power supply (TC-47)	2020/04/06	2020/04/06 17:30:32	passed
<b>SPEC-32</b> Expired air	<b>XTC-50</b> Expired air (TC-29)	2020/04/05	2020/04/05 07:26:32	passed
<b>SPEC-44</b> Non-conductive exterior casing	<b>XTC-52</b> Non-conductive exterior casing (TC-74)	2020/04/05	2020/04/05 08:05:32	passed
<b>SPEC-45</b> Protective equipment	<b>XTC-119</b> Usability with protective equipment (TC-41)	2020/04/24	2020/04/24 09:40:03	passed
<b>SPEC-23</b> Protected off switch	<b>XTC-99</b> Off switch protection (TC-21)	2020/04/09	2020/04/09 18:08:37	passed
<b>SPEC-22</b> The O2 and air input connectors	<b>XTC-85</b> O2 and air input connectors (TC-20)	2020/04/07	2020/04/07 09:05:22	passed
<b>SPEC-18</b> Battery use warning	<b>XTC-129</b> Back up battery warning (TC-69)	2020/04/24	2020/04/24 11:07:59	failed

## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 8 / 75

Item	Tested By	Execution Date	Last Modification	Test Result
<b>SPEC-5 Pressure relief valve</b>	<b>XTC-106 Fail-safe valve (TC-67)</b>	2020/04/10	2020/04/15 11:22:04	passed
<b>SPEC-4 Calibrate sensors during assembly</b>	<b>XTC-18 Calibrate sensors during assembly (TC-66)</b>	2020/04/01	2020/04/01 14:20:00	passed with deviation
<b>SPEC-2 Robustness internal components</b>	<b>XTC-120 Internal components tightly attached (TC-4)</b>	2020/04/24	2020/04/24 09:46:59	passed with deviation
<b>SPEC-1 Ingress of water</b>	<b>XTC-98 Splash proof (TC-3)</b>	2020/04/09	2020/04/09 18:06:49	passed
<b>SPEC-162 Installation Verification</b>	<b>XTC-132 Installation Verification (TC-174)</b>	2020/04/24	2020/04/25 10:37:38	passed
<b>SPEC-152 IFU: regulatory requirements</b>	<b>XTC-135 IFU: regulatory requirements (TC-165)</b>	2020/04/29	2020/04/29 09:45:44	passed with deviation
<b>SPEC-47 IFU: intended use</b>	<b>XTC-35 Clear explanation for intended use in IFU (TC-134)</b>	2020/04/02	2020/04/02 09:17:10	passed
<b>SPEC-13 IFU: Device not for use during transport</b>	<b>XTC-32 IFU: Not for use during transport (TC-12)</b>	2020/04/02	2020/04/02 08:58:38	passed
<b>SPEC-14 IFU: Check device before use</b>	<b>XTC-33 IFU: Check device before use (TC-13)</b>	2020/04/02	2020/04/02 09:03:16	passed
<b>SPEC-98 IFU: Do not use when damaged</b>	<b>XTC-31 IFU: Unplug device when damaged (TC-114)</b>	2020/04/02	2020/04/02 09:00:06	passed
<b>SPEC-117 IFU: turn off device before unplugging</b>	<b>XTC-36 IFU: Device turned off (TC-117)</b>	2020/04/02	2020/04/02 09:18:56	passed
<b>SPEC-134 IFU: ventilator settings</b>	<b>XTC-39 IFU: Upper en lower limit of the ventilator settings (TC-135)</b>	2020/04/02	2020/04/02 09:24:30	passed
<b>SPEC-135 IFU: Explanation of the alarms</b>	<b>XTC-41 IFU: explanation of the alarms (TC-136)</b>	2020/04/02	2020/04/02 09:29:07	passed
<b>SPEC-137 IFU: HEPA filter</b>	<b>XTC-44 IFU: Instructions about the use of the HEPA-filter (TC-137)</b>	2020/04/02	2020/04/02 09:37:05	passed
<b>SPEC-51 IFU: Instructions for correct and safe maintenance</b>	<b>XTC-80 IFU: Maintenance (TC-44)</b>	2020/04/07	2020/04/07 08:02:10	passed



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 9 / 75

Item	Tested By	Execution Date	Last Modification	Test Result
<b>SPEC-138</b> IFU: alarm boundaries	<b>XTC-45</b> IFU: How to set the upper and lower boundaries of the alarms (TC-138)	2020/04/02	2020/04/02 09:40:54	passed with deviation
<b>SPEC-17</b> Training and clear IFU for the users	<b>XTC-67</b> IFU: trained user (TC-75)	2020/04/06	2020/04/06 10:01:52	passed
	<b>XTC-110</b> Usability test (TC-65)	2020/04/10	2020/04/24 09:47:51	documentation
<b>SPEC-49</b> IFU: checked by experts	<b>XTC-136</b> IFU Review (TC-43)	2020/04/29	2020/04/29 09:52:20	passed
<b>SPEC-151</b> Label: regulatory requirements	<b>XTC-137</b> Label: regulatory requirements (TC-164)	2020/04/28	2020/04/29 09:54:13	passed
<b>SPEC-64</b> Emergency use sticker	<b>XTC-13</b> Labeling: Warning Sticker (TC-105)	2020/03/31	2020/03/31 16:01:10	passed
<b>SPEC-115</b> Emergency number on label	<b>XTC-12</b> Sticker with emergency number on device (TC-113)	2020/03/31	2020/03/31 15:58:50	passed
<b>SPEC-133</b> HEPA filter warning sticker	<b>XTC-126</b> HEPA filter warning sticker (TC-173)	2020/04/24	2020/04/24 10:34:53	passed
<b>SPEC-61</b> Do not push warning	<b>XTC-11</b> Do not push warning (TC-129)	2020/03/31	2020/03/31 15:58:06	passed
<b>SPEC-154</b> Fire warning sticker	<b>XTC-108</b> INSPECT: Information about fire hazard on the label (TC-166)	2020/04/20	2020/04/20 10:42:28	passed
<b>SPEC-148</b> Device label	<b>XTC-121</b> Visibility of label (TC-160)	2020/04/24	2020/04/24 09:49:13	passed
<b>SPEC-130</b> Intuitive for use - training	<b>XTC-78</b> Intuitive for use - training (TC-131)	2020/04/06	2020/04/06 19:01:40	passed
<b>SPEC-48</b> Vocabulary and Semantics	<b>XTC-138</b> Vocabulary and Semantics (TC-71)	2020/04/29	2020/04/29 09:55:30	passed
<b>SPEC-167</b> Open Source	<b>XTC-133</b> Open Source (TC-175)	2020/04/24	2020/04/24 14:05:27	passed
<b>SPEC-124</b> Pre-Set Controls	<b>XTC-83</b> Pre-Set Controls (TC-125)	2020/04/07	2020/04/07 09:00:46	passed
<b>SPEC-121</b> Battery Housing	<b>XTC-104</b> Battery Housing (TC-122)	2020/04/13	2020/04/13 12:21:08	passed
<b>SPEC-112</b> Incorrect output	<b>XTC-84</b> Incorrect output (TC-98)	2020/04/07	2020/04/07 09:02:13	passed
<b>SPEC-111</b> Indication of parameters relevant to safety	<b>XTC-88</b> Indication of parameters relevant to safety (TC-97)	2020/04/07	2020/04/07 09:22:04	passed



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 10 / 75

Item	Tested By	Execution Date	Last Modification	Test Result
<b>SPEC-110 Intentional exceeding of safety limits</b>	<b>XTC-100 Intentional exceeding of safety limits (TC-96)</b>	2020/04/10	2020/04/10 08:07:54	passed
<b>SPEC-109 Alarm systems</b>	<b>XTC-24 Alarmsystems (TC-95)</b>	2020/04/01	2020/04/01 15:26:10	passed
<b>SPEC-108 Usability</b>	<b>XTC-89 Usability (TC-94)</b>	2020/04/07	2020/04/07 09:28:25	passed with deviation
<b>SPEC-107 Accuracy of controls and instruments</b>	<b>XTC-90 Accuracy of controls and instruments (TC-93)</b>	2020/04/07	2020/04/07 11:21:34	passed
<b>SPEC-105 Compatibility with substances</b>	<b>XTC-8 Compatibility with substances (TC-90)</b>	2020/03/30	2020/03/30 08:22:16	passed
<b>SPEC-102 Flammable agent</b>	<b>XTC-7 ME intended for use in conjunction with flammable agent (TC-88)</b>	2020/03/30	2020/03/30 08:21:17	passed
<b>SPEC-77 Arrangements of controls and indicators</b>	<b>XTC-105 Arrangements of controls and indicators (TC-59)</b>	2020/04/13	2020/04/13 12:24:03	passed
<b>SPEC-76 Electromagnetic compatibility</b>	<b>XTC-122 Electromagnetic compatibility (TC-58)</b>	2020/04/24	2020/04/24 09:50:16	passed

## TEST REPORTS

<b>XTC-32 IFU: Not for use during transport (TC-12)</b>
INSPECTION Warning in IFU that device is not to be used during transport
1
Version 1.0 of the IFU.
Legal OperationAIR
2020/04/02
passed
Page 19 under section 10.



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 11 / 75

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect Instructions For Use	Contains a warning that device is not to be used during transport	passed	

### XTC-33 IFU: Check device before use (TC-13)

#### INSPECTION

IFU contains information on what inspection must be done before use

1

Version 1.0 of the IFU

Legal OperationAIR

2020/04/02

passed

Page 13 under section 5 subsection 1.

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect instructions for use	contains information on inspections to be done before using the device	passed	

### XTC-34 Tidal volume range (TC-23)

#### TEST

Confirm that the device is able to provide tidal volumes between 300 and 700 mL

1

2020/04/02

passed

1. tV: >> 700mL (888mL)
2. tV: << 300 mL (15mL)



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 12 / 75

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Connect device to MI test lung			
2	Set lung compliance to 0.1 L/cmH2O			Compliance = 0.4 L/cmH2O
3	Ventilate with PPLAT 50 cm H2O, PEEP 5 cm H2O, freq 10/min			
4	Read out tidal volume on MI lung	tidal volume >= 700 mL		
5	Set lung compliance to 0.01 L/cmH2O			Compliance = 0.4 L/cmH2O
6	Ventilate with PPLAT 10 cm H2O, PEEP 5 cm H2O, freq 10/min			
7	Read out tidal volume on MI lung	tidal volume <= 300 mL		

### XTC-43 PEEP value (TC-30)

TEST

Confirm that device is able to provide PEEP of at least 15 cm H2O

1

Prototype

Quality Control OperationAIR

2020/04/06

passed

PEEPlung 19.51 cm H2O

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Connect device to MI test lung			
2	Ventilate with PEEP 20 cm H2O			
3	Read out PEEP in MI test lung	PEEP > 15 cm H2O	passed	

### XTC-2 Visualized measurements (TC-31)

INSPECTION

Flow, Pressure and Volume are visualized on the touchscreen



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 13 / 75

1.0				
Prototype 1.0				
Quality Control OperationAIR				
2020/03/27				
passed				
Pass				
Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect user interface	Graphs and numbers showing values for Flow, Volume and Pressure are depicted	passed	

### XTC-3 Touchscreen gloves (TC-41)

DEMONSTRATION				
Confirm that the user interface touch screen works when the user wears gloves				
1.0				
Isolated component				
Quality Control OperationAIR				
2020/03/26				
passed				
Works perfect with a single glove. There may be an almost imperceivable latency when wearing two gloves				
Step	Action Result	Expected Result	Passed/Failed	Comment
1	Put on two layers of protective gloves			
2	Adjust some settings on the user interface	Touchscreen responds to touching	passed	



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 14 / 75

### XTC-19 Sound test (TC-49)

The sound emitted by the device can not exceed 80 dBA. Sound level meters used in the measurement conform to IEC 61672-1 and IEC 61672-2. The test room is semi-reverberant with a hard reflecting floor. The distance between any wall or other object and the surface of the device is not less than 3 m.

1

Prototype

QA OperationAIR

2020/04/06

passed

sound level peaks at 56 dBA

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Place device in room with hard reflecting floor, at least 3 m distance to every wall			
2	Turn device on maximum pressure and flow (to find the worst-case scenario)			
3	Measure dBA at 50 cm from the device	does not exceed 80 dBA	passed	

### XTC-18 Calibrate sensors during assembly (TC-66)

Inspection

1

Documentation

Quality Control OperationAIR

2020/04/01

passed with deviation



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 15 / 75

Installation verification protocol contains a check if calibration certificates are available for the relevant sensors (pressure and flow)

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect Installation Verification protocol	Sensor calibration is mentioned	passed	

### XTC-14 HME filter (TC-70)

#### INSPECTION

Show that the device can be equipped with an ISO 9360-1:2000 certified HME filter

1

Prototype

Quality Control OperationAIR

2020/03/31

passed

HME filter is placed at patient airway side. The standard hospital equipment is compatible with standard HME filters and this device

Step	Action Result	Expected Result	Passed/Failed	Comment
1	inspection		passed	Included documentation in inspection

### XTC-30 Error margins: Pressure loss (TC-76)

#### ANALYSIS

Maximum loss of pressure between device and patient lungs is 10 cm H<sub>2</sub>O

2

2020/04/02

passed

1. PIP: 11.7 cmH<sub>2</sub>O, Difference: -3.3
  - o PEEP: 4.1 cmH<sub>2</sub>O, Difference: -0.9



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 16 / 75

- 2. PIP: 26.6 cmH2O, Difference: -3.4
  - o PEEP: 19.3 cmH2O, Difference -0.7
- 3. PIP: 41.7 cmH2O, Difference: -3.3
  - o PEEP: 34.5 cmH2O, Difference -0.5
- 4. PIP: 56.6 cmH2O, Difference: -3.4
  - o PEEP: 49.2 cmH2O, Difference: -0.8

**Kalibreren!!!**

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Connect device to MI test lung			
2	Ventilate with PPLAT 15, PEEP 5 cm H2O, I:E 0.5, Freq 15, FiO2 21%			
3	Read out Plateau pressure on MI test lung			
4	Increase PPLAT to 60 in 15 cm H2O interval and the peep to 50 in 15 cm H2O interval and read out Plateau pressure on MI lung each time			
5	Calculate deviation between set and measured values	Deviation < 10 cm H2O for each measurement		

### XTC-25 Error margins: Breating rate (TC-78)

**ANALYSIS**

confirm that maximum error between measured and set breating rates is 1/min

2

Quality Control OperationAIR

2020/04/06

passed

pass

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Connect device to MI test lung			
2	Ventilate with PPLAT 25 cm H2O, PEEP 15 cm H2O, I:E 0.5, FiO2 21%			
3	Set Freq to 10/min			



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 17 / 75

4	Read out breathing rate on MI test lung			
5	Increase freq to 30/min in intervals of 10/min and read out breathing rate on MI test lung each time			
6	Calculate deviation between set and measured values	deviation < 1/min	passed	

### XTC-26 Error margins: FiO2 (TC-79)

#### ANALYSIS

Confirm that maximum error margin between measured and set FiO2 is 5%

2

Prototype

Quality Control OperationAIR

2020/04/02

passed

#### FiO2 Setting | Measurement | Deviation

21% | 21% | 0%

40% | 41% | 1%

60% | 60% | 0%

80% | 79% | 1%

100% | 99% | 1%

Mean deviation: 1%

Mean settle time 3:30

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Connect device to MI test lung			
2	Connect external O2 sensor at test lung			
3	Ventilate with PPLAT 25 cm H2O, PEEP 15 cm H2O, I:E 0.5, Freq 15			
4	Set FiO2 to 21%			
5	Read out O2 on external sensor			Additionally measured time until O2 sensor settles



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 18 / 75

6	Increase FiO2 to 100% in 20% intervals and read out O2 on external sensor each time			
7	Calculate deviation between set and measured values	Deviation < 5%	passed	Mean deviation 1%

### XTC-27 Error Margins: I:E ratio (TC-80)

#### ANALYSIS

Confirm that maximum error between set and measured I:E ratio is 0.05

2

2020/04/02

failed

1. I:E = 2.35 , Difference: 0.35
2. I:E = 3.45 , Difference:0.45 (alternates between 3.32 en 3.48)
3. I:E = 4.75 , Difference:0.75 (alternates between 4.70 en 4.82)

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Connect device to MI test lung			
2	Ventilate with PPLAT 25 cmH2O, PEEP 15 cmH2O, Freq 15, FiO2 21%			
3	Set I:E to 0.5			
4	Read out I:E on MI test lung	deviation < 0.05		
5	Set I:E to 0.33			
6	Read out I:E on MI test lung	deviation < 0.05		
7	Set I:E to 0.25			
8	Read out I:E on MI test lung	deviation < 0.05		

### XTC-28 Error margins: Tidal Volume (TC-81)

#### ANALYSIS

Confirm that maximum error margin between internally and externally measured tidal volume is 5%

2



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 19 / 75

Prototype

Quality Control OperationAIR

2020/04/02

passed with deviation

1. tV\_testlung 77 mL: , tV\_device: 82, deviation 6.0%
  2. tV\_testlung 364 , tV\_device 339, deviation 7.3%
  3. tV\_testlung 811 , tV\_device 821, deviation 1.2%
  4. tV\_testlung 885, tV\_device 860, deviation 2.9%
  5. tV\_testlung : , tV\_device Not measured because maximum volume of test lung was exceeded
- Deviation seems to be a constant rather than percentual. Averages out < 5%

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Connect device to MI test lung			
2	Ventilate with PEEP 5cm H2O, I:E 0.5, FiO2 21%			
3	Set PPLAT to 10 cm H2O, freq 30/min, lung compliance 0.01 L/cmH2O			
4	Read out tidal volume on MI test lung AND on internal sensor			
5	Increase PPLAT to 50 in 10 cm H2O intervals; decrease freq to 5/min in 5/min intervals; increase lung compliance to 0.1 in 0.02 L/cmH2O intervals. Read out tidal volume for each measurement on MI test lung AND on internal sensor			
6	Calculate deviations between values	deviations < 5%	passed	Mean deviation < 5%

**XTC-7 ME intended for use in conjunction with flammable agent (TC-88)**

INSPECTION

Adress risk of fire in risk management file

1

RMF 30/03/2020



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 20 / 75

Quality Control OperationAIR				
2020/03/30				
passed				
Risks of flammable agent (oxygen) addressed in RMF				
Step	Action Result	Expected Result	Passed/Failed	Comment

<b>XTC-8 Compatibility with substances (TC-90)</b>				
Address risks associated with compatibility with substances used with the medical device (oxygen) in risk management file				
1				
RMF 30/03/2020				
Quality Control OperationAIR				
2020/03/30				
passed				
Risk of oxygen addressed in RMF				
Step	Action Result	Expected Result	Passed/Failed	Comment

<b>XTC-24 Alarmsystems (TC-95)</b>				
Address need for alarm systems by means of risk control and address the risks associated with the operation or failure of the alarm system in Risk management File.				
1				
documentation				
Quality Control OperationAIR				



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 21 / 75

2020/04/01				
passed				
Adressed in Risk management file				
Step	Action Result	Expected Result	Passed/Failed	Comment

### XTC-13 Labeling: Warning Sticker (TC-105)

INSPECTION				
Show that the label on the device indicates that the device is only to be used in emergency situations on COVID-19 patients				
1				
isolated label design				
Quality Control OperationAIR				
2020/03/31				
passed				
Has text on label indicating the intended use for covid-19 patients				
Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect device	has warning sticker	passed	

### XTC-9 230 V circuit outside casing (TC-112)

INSPECTION				
Show that the 230 V adapter and circuitry are not in the same encasing as the airflow circuits				
1				
Prototype				
Quality Control OperationAIR				



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 22 / 75

2020/03/30				
passed				
<p>PASS 230V adapter and battery are placed in a separate encasing at distance of the main encasing that contains the oxygen circuitry</p>				
Step	Action Result	Expected Result	Passed/Failed	Comment

<b>XTC-12 Sticker with emergency number on device (TC-113)</b>				
INSPECTION				
Inspect Device for labelling. Should name the number to call in case of emergency.				
1				
isolated label design				
Quality Control OperationAIR				
2020/03/31				
passed				
Has telephone number on label				
Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect Labelling		passed	

<b>XTC-31 IFU: Unplug device when damaged (TC-114)</b>				
INSPECTION				
Show that IFU mentions to unplug device when damaged				
1				
Version 1.0 of the IFU.				



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 23 / 75

Legal OperationAIR

2020/04/02

passed

Page 19 under section 10.

Step	Action Result	Expected Result	Passed/Failed	Comment
1	inspect IFU		passed	

### XTC-36 IFU: Device turned off (TC-117)

INSPECT

Warning in IFU that the device should be turned off

1

Version 1.0 of the IFU

Legal OperationAIR

2020/04/02

passed

Page 14 under section 6 subsection 2.

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect IFU		passed	

### XTC-38 Peak-plateau difference (TC-121)

TEST

Show that the difference between peak and plateau pressure is < 2 cm H<sub>2</sub>O at default ventilation settings

1

Prototype



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 24 / 75

### Quality Control OperationAIR

2020/04/02

passed

PIP: 23.2

PEEP: 9.0

initial analysis PPLAT: 9.0

After final analysis: Peak-plateau difference mean 1.08 cm H2O over 1 minute

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Connect device to MI test lung			
2	Set compliance to 0.03 L/cmH2O			
3	Ventilate with PPLAT 25, PEEP 10, Freq 10, I:E 0.5, FiO2 21%			
4	Read out peak lung pressure and pplat on mi lung			
5	Calculate difference between measured PPLAT and PIP	difference < 2 cm H2O	passed	1.08

### XTC-11 Do not push warning (TC-129)

#### DEMONSTRATION

Show that the device has a sticker that says to not push on the device

1

Isolated label design

### Quality Control OperationAIR

2020/03/31

passed

Has iso 7010-P017 icon on label

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Demonstration		passed	



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 25 / 75

### XTC-35 Clear explanation for intended use in IFU (TC-134)

INSPECT

Check if the intended use is documented in the IFU

1

Version 1.0 of the IFU

Legal OperationAIR

2020/04/02

passed

Page 5 under section 1 subsection 1.

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect the IFU	It is documented	passed	

### XTC-46 PEEP stability (TC-133)

TEST

Confirm that the set PEEP level is maintained at all times

1

Prototype

Quality Control OperationAIR

2020/04/06

passed

PEEP 13.93 SD 0.12 over 1 minute

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Connect device to test lung			



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 26 / 75

2	Ventilate with PPLAT 25, PEEP 15, Freq 15			
3	Ventilate for 1 minute			
4	Determine PEEP levels for each breath	deviation between PEEP levels < 1 cm H2O	passed	

### XTC-39 IFU: Upper en lower limit of the ventilator settings (TC-135)

INSPECT

Check if the upper and lower limits of the ventilator are explained correctly. Check whether the IFU explains how to set your ventilator parameter.

1

Version 1.0 of the IFU

Legal OperationAIR

2020/04/02

passed

Page 12 under section 4.

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect IFU	Correctly documented way to set your settings	passed	

### XTC-41 IFU: explanation of the alarms (TC-136)

INSPECT

Check if the alarms, their causes and how to deal with them are clearly explained in the IFU.

1

Version 1.0 of the IFU

Legal OperationAIR

2020/04/02

passed



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 27 / 75

Page 15 under section 7 and pages 20 & 21 under section 11.

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect IFU	Alarms are explained	passed	

### **XTC-44 IFU: Instructions about the use of the HEPA-filter (TC-137)**

INSPECT

Check whether correct usage of the HEPA-filter is explained in the IFU

1

Version 1.1 of the IFU

Legal OperationAIR

2020/04/02

passed

Page 13 under section 5.

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect IFU	The instructions about the HEPA filter are correctly documented	passed	

### **XTC-45 IFU: How to set the upper and lower boundaries of the alarms (TC-138)**

INSPECT

Check whether instructions about how to set the alarm limit is documented

1

Version 1.0 of the IFU

Legal OperationAIR

2020/04/02



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 28 / 75

passed with deviation

Page 8 under section 3 subsection 2.

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect IFU	Alarm instructions are well documented	passed	Still has to be supplemented after completion of the GUI. But basics are there. Screenshots have to be implemented.

### XTC-47 Patient leakage current (TC-53)

The values for patient leakage must not exceed:

- Alternating current maximally 500 micro Ampere, direct current maximally 50 micro Ampere in normal condition
- Alternating current maximally 1000 micro Ampere, direct current maximally 100 micro Ampere in single fault condition

1

not started

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Measure leakage current at patient level	A current < 500 uA is measured		

### XTC-48 Touch current (TC-52)

Touch current from or between parts of the medical device within the patient environment must not exceed 100  $\mu$ A. Leakage current from accessible outer surfaces of the equipment is also considered to be touch current.

1

not started

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Measure touch current for the different parts and outer surface of the medical device	Touch current is maximally 100 $\mu$ A.		

### XTC-49 Pressure Rise Time (TC-120)



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 29 / 75

### TEST

Show that pressure rise time to peak at default ventilation settings is lower than 200 ms

1

Prototype

Quality Control OperationAIR

2020/04/05

**failed**

10%-90% pressure rise time as measured at the input of the device was ca. 300 ms.  
When measured at the output, the pressure rise time was ... ms.

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Connect device to MI test lung			
2	Set lung compliance to 0.03 L/cmH2O			
3	Ventilate with PPLAT 25, PEEP 10, Freq 10, I:E 0.5			And resistance of 5
4	Read out pressure rise time	< 200 ms	<b>failed</b>	

### **XTC-50 Expired air (TC-29)**

#### DEMONSTRATION

Show that the device has a check valve to block expiratory air from entering the device

terugslagklep testen op 80

1

Prototype

Quality Control OperationAIR

2020/04/05

passed



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 30 / 75

Check valve is present

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Show check valve		passed	

### XTC-51 Breaks (TC-103)

Inspection  
Show that the device has breaks on its wheels

1

Prototype

Quality Control OperationAIR

2020/04/05

passed

Inspection of prototype and documentation. Device has 4 wheels with breaks for the two on the front side.

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect wheels	wheels have breaks	passed	

### XTC-52 Non-conductive exterior casing (TC-74)

DEMONSTRATION  
Show that the external casing is made of a non-conductive material and that conductive parts are grounded

1

Prototype

Quality Control OperationAIR

2020/04/05

passed



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 31 / 75

Inspect prototype and documentation. Casing is made of Polyurethane which is an insulator. The stainless steel bottomplate is grounded

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect design schematic			
2	Inspect casing datasheet			

### XTC-53 Protective earth connections (TC-56)

PROTECTIVE EARTH CONNECTIONS shall be made so that the removal of any single item of equipment in the ME SYSTEM will not interrupt the protective earthing of any other part of the ME SYSTEM, without at the same time disconnecting the electrical supply to that part. Additional PROTECTIVE EARTH CONDUCTORS shall only be detachable by use of a TOOL.

1

All components are grounded via the bottomplate. Removal of one component will therefore not lead to interruption of protective earthing. Removal of the bottomplate could disrupt grounding, but this is impossible without completely disassembling the device, so this is considered acceptable.  
The additional protective earth conductor can be easily removed

Quality Control OperationAIR

2020/04/05

**failed**

Failed. Additional protective earth is easily removed without a tool

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect design schematics			

### XTC-56 Flow sensor (TC-24)

INSPECTION

Show that the device is equipped with a flow sensor

1

Design schematic



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 32 / 75

Quality Control OperationAIR

2020/04/05

passed

Device has two flow sensors, one expiratory and one inspiratory

Step	Action Result	Expected Result	Passed/Failed	Comment
------	---------------	-----------------	---------------	---------

### XTC-57 FiO2 setting (TC-34)

DEMONSTRATION

Confirm that the FiO2 can be set between 21% and 100% in intervals of 5%

1

Prototype

Quality Control OperationAIR

2020/04/05

passed

User interface has option to set FiO2

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Demonstrate FiO2 setting		passed	
2				

### XTC-58 External exhaust outlet oxygen rich environment (TC-86)

INSPECTION

External exhaust outlets must not be located so that risk of ignition occurs because of any electrical component mounted on the outside of the medical device.

1

design schematic and prototype



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 33 / 75

Quality Control OperationAIR

2020/04/05

passed

All exhausts are placed at the bottom of the device, where no electrical equipment is placed

Step	Action Result	Expected Result	Passed/Failed	Comment
------	---------------	-----------------	---------------	---------

### XTC-60 PEEP settings (TC-32)

DEMONSTRATION

Confirm that the PEEP can be set between 5 and 70 cm H2O and is adjustable per 5 cm H2O

1

2020/04/05

passed

PEEP can be set between 5 and 70 cmH2O and is adjustable per 5 cmH2O

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Demonstrate PEEP settings			
2				

### XTC-61 Fail-safe valve (TC-67)

TEST

Confirm that the fail-safe valve opens and releases pressure when pressure exceeds 70 cm H2O

1

2020/04/05

passed

Pressure reaches 100 cmH2O at first, but drops to 70 as soon as pressure reaches Fail Safe Valve.



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 34 / 75

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Attach pressure sensor to airflow output			
2	Open proportional valve to put 100 cm H2O on system			Via PC because settings on GUI are limited
3	Read out pressure sensor	Pressure has not exceeded 70 cm H2O (+- 1 cm H2O)		Expected result: When ventilation stops, pressure instantly drops to 70 cmH2O

### XTC-63 Leakage current testing (TC-54)

#### INSPECTION

Factory acceptance test protocol states that each manufactured device is checked for leakage current and shall be < 500 uA.

2

IV protocol v 1.1

Quality Control OperationAIR

2020/04/10

passed

Stated leakage current testing in installation verification protocol

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect factory acceptance test protocol		passed	

### XTC-64 Off switch protection (TC-21)

#### DEMONSTRATION

Show that the off switch is positioned so that it is protected from accidental switching

1



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 35 / 75

Prototype

Quality Control OperationAIR

2020/04/05

passed

Device has an off switch on the user interface. When pressed, the user is prompted to confirm switching off the device. The power adapter has a button to turn of charging the battery, but this is located relatively protected

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Find location of off switch, try to bump into it		passed	

### XTC-65 Inspiratory Hold (TC-156)

Medical device must be equipped with a button that will stop respiration at the end of inspiration when pressed, while measuring the pressure in the system. When the user lets go of the button, respiration must instantly start again

1

2020/04/10

passed with deviation

Inspiratory hold is started by a button and stopped by another button, or when exceeding a set time limit

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Press inspiratory hold button	Respiration stops at end of inspiration, Pressure over the system is measured	passed	
2	Let go of button	Respiration is restored	passed	after 7 seconds

### XTC-66 IFU: Use breaks when needed (TC-104)

INSPECTION

Confirm that the instructions mention that the breaks need to be used when the device is not being moved

1



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 36 / 75

Version 1.2 of the IFU				
Legal OperationAIR				
2020/04/06				
passed				
Page 16 under section 10.				
Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect IFU	Mention to use breaks	passed	

### XTC-67 IFU: trained user (TC-75)

Warning in IFU that the device may only be used by trained users				
1				
Version 1.2 of the IFU				
Legal OperationAIR				
2020/04/06				
passed				
Page 1 on title page				
Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect IFU		passed	

### XTC-68 Tipping test (TC-109)

TEST Confirm that the device does not overbalance when placed on a 10 degree incline surface				
1				



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 37 / 75

Prototype

Quality Control OperationAIR

2020/04/06

passed

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Put device on hard, flat horizontal, movable surface			
2	Put device on breaks			
3	Tilt surface to 10 degree angle to horizontal	device does not fall over	passed	

### **XTC-69 I:E ratio (TC-148)**

Inspection

The inspiratory:expiratory ratio is visible and adjustable on GUI.

1

Prototype

Quality Control OperationAIR

2020/04/24

passed

I:E can be set to:

1:0.5; 1:1; 1:1.5; 1:2; 1:2.5; 1:3

Step	Action Result	Expected Result	Passed/Failed	Comment
------	---------------	-----------------	---------------	---------

### **XTC-70 Inspiratory flow (TC-37)**

TEST

Confirm that maximum inspiratory flow > 1.5 L/s



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 38 / 75

1				
prototype				
Quality Control OperationAIR				
2020/04/06				
passed				
Flow = 105 L/min				
Step	Action Result	Expected Result	Passed/Failed	Comment
1	Connect device to MI test lung with flow sensor			
2	Set MI lung compliance to 0.1 L/cmH2O			
3	Ventilate with PPLAT 50, PEEP 5, I:E 0.5, freq 10/min, FiO2 21%			
4	Read out peak inspiratory flow	> 90 L/min	passed	
5				

### XTC-74 Power supply (TC-47)

#### DEMONSTRATION

Show that the device has a IEC 60601-1 certified mains supply adapter for connection to a 230V AC supply mains

1				
2020/04/06				
passed				
Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect IFU	Supply mains connection to 230V AC mentioned		
2	Inspect datasheet of supply mains adapter	Has IEC 60601-1 certification		



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 39 / 75

### XTC-75 Power Supply Connection (TC-123)

**INSPECTION**

Power supply cannot be incorrectly connected or replaced

1

2020/04/06

passed

Connections of Power outlet to battery and battery to device ar not interchangeable.

Step	Action Result	Expected Result	Passed/Failed	Comment
------	---------------	-----------------	---------------	---------

### XTC-76 Protection of conductors (TC-57)

Conductors that connect different items of equipment within an ME SYSTEM shall be protected against mechanical damage.

1

Design schematics, prototype

Quality Control OperationAIR

2020/04/06

passed

All conductive parts are placed inside either one of the casings, or have a strain-relief

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect, try to damage mechanically	Conductor stays intact	passed	

### XTC-77 Compatibility with standard hospital equipment (TC-39)

**DEMONSTRATION**

Show that connectors and components are compatible with standard equipment based on the relevant standards



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 40 / 75

1				
Prototype				
Quality Control OperationAIR				
2020/04/06				
passed				
Airflow tubing connectors are compatible with the standard 22 mm tubing, as specified in EN-ISO-5361:2016 O2 and pressurized air input connectors are compatible with the standard male NIST connectors as specified in ISO 18082:2014				
Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect datasheets		passed	

### XTC-78 Intuitive for use - training (TC-131)

Must not require more than 30minutes training for a doctor with some experience of ventilator use.				
1				
Documentation				
Quality Control OperationAIR				
2020/04/06				
passed				
Training materials indicate 20 minute training for users				
Step	Action Result	Expected Result	Passed/Failed	Comment

### XTC-79 Long-term reliability (TC-38)

TEST Reliability test with 12 hours as initial investigation into the reliability of the device Additionally:				
---	--	--	--	--

- Measure temperature on several moments in time
- Post-analysis: pplat and peep stay constant. moving average coefficient = 0

1

Prototype

Quality Control OperationAIR

2020/04/06

passed

**Log**

22:00 Start measurement  
 22:20 Temperature inside casing bottomplate 27.0, expiratory valve 50.1, Pi 47.2  
 22:59 stopped ventilation to add pulmotech pressure catheter to measurement setup  
 23:53 ambient: 28, Valve: 53, GUI/pie: 51 degrees Celsius  
 00:41 ambient 26.7, Valve 50.6, Pi 45,6 degrees. Ventilation parameters perfect  
 01:10 ambient: 29, Valve: 53, Pi: 51, variables are amazing  
 2:04 ambient 28, Valve 51, Pi 49  
 2:29 ambient 27.9, valve 51.4, Pi 47.5. Ventilation parameters ok, tidal volume deviation +- 50 ml  
 3:02 ambient 17.9 Valve 52.6 Pi 50.8  
 3:05 Deken er omheen (per ongeluk op stop knop gedrukt!!!!!! --> Moet eigenlijk niet kunnen natuurlijk)  
 3:16 Deken eraf Ambient 30, Valve 52, GUI 51  
 4:17 Ambient 28, Valve 53, GUI 52  
 5:00 Ambient 28, Valve 53, GUI 51 Vt deviation c. 50 cmH2O  
 5:36 Ambient 26.9, valve 50.2, Gui 49.3  
 6:32 Ambient 26.9, valve 50.2, Gui 48.9  
 8:21 Ambient 27.6, valve 49.3, gui 44.9  
 8:57 Ambient 28.0, valve 50.0, gui 49,6  
 9:42 Ambient 26.7, valve 49.6, gui 47.5  
 10:00 Stopped measurement

**Data analysis**

Two points where device stops ventilating, explained by manual stopping, see log.  
 Trend in Peak pressure and PEEP over time is stable (directional coefficient 1.00,  $R^2 < 1 * 10^{-10}$ )

**Conclusion**

Device functions as expected, no unexplained interruptions. Temperature of components did not exceed 53 degrees celsius.



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 42 / 75

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Connect device to test lung			
2	Ventilate with PPLAT 35 cm H2O, PEEP 10 cm H2O, Freq 25/min, FiO2 21%, I:E 0.5			
3	Keep ventilating for at least 12 hours	Device stays active throughout the whole time	passed	
4	Measure internal temperature at 4 points in time			
5	Analyse temperature	Temperature inside casing does not exceed 60 degrees celsius	passed	

### XTC-80 IFU: Maintenance (TC-44)

#### DEMONSTRATION

Show that the IFU states maintenance instructions

1

Version 1.3 of the IFU

Legal OperationAIR

2020/04/07

passed

Page 15 and 16 under section 8.

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect IFU		passed	

### XTC-81 Specification of expected durability (TC-130)

Specify expected durability in documentation.

1

All documentations are tested (IFU, Technical manual), Onderzoeksprotocol, IMDD, RMF.

Legal OperationAIR



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 43 / 75

2020/04/07

passed with deviation

All documentation mention the intended use of a minimum of four weeks. All parts are more durable given correct usage. Because we cannot test durability, this is the most we can do.

Step	Action Result	Expected Result	Passed/Failed	Comment
------	---------------	-----------------	---------------	---------

### XTC-82 Back up battery (TC-10)

TEST

Confirm that the UPS continues device functionality in case of power failure

1

Prototype

Quality Control OperationAIR

2020/04/07

passed

Device keeps functioning as expected without interruption when disconnected from mains power supply. Ventilation parameters do not change.  
Battery keeps device functional for 36 minutes 16 seconds

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Connect device to MI test lung			
2	Ventilate with PPLAT 35 cm H <sub>2</sub> O, PEEP 5 cm H <sub>2</sub> O, freq 30/min			
3	Read out Tidal volume and plateau pressure on test lung			
4	Cut off the main power supply	Device keeps ventilating	passed	
5	Read out tidal volume and plateau pressure on test lung	Values should stay within 1% deviation from values read out at step 3	passed	



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 44 / 75

6	Time untill battery fails	> 30 minutes	passed	
---	---------------------------	--------------	--------	--

### XTC-83 Pre-Set Controls (TC-125)

Risks associated with pre-set controls must be addressed in Risk Management File

1

RMF of today (see date).

Legal OperationAIR

2020/04/07

passed

Documented in risk 39 and 40

Step	Action Result	Expected Result	Passed/Failed	Comment
------	---------------	-----------------	---------------	---------

### XTC-84 Incorrect output (TC-98)

Risks associated with incorrect output must be addressed in risk management file.

1

RMF of today

Legal OperationAIR

2020/04/07

passed

Risks under folder H02

Step	Action Result	Expected Result	Passed/Failed	Comment
------	---------------	-----------------	---------------	---------

### XTC-85 O2 and air input connectors (TC-20)

DEMONSTRATION



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 45 / 75

show the different inlets and show that the O2 cannot be connected to the air inlet and that air cannot be connected to the O2 inlet

1

Prototype

Quality Control OperationAIR

2020/04/07

passed

Connectors cannot be interchanged

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Attempt to connect O2 to air inlet	Will not connect	passed	
2	Attempt to connect air to the O2 inlet	will not connect	passed	

### XTC-86 Oxygen alarm (TC-9)

TEST

Confirm that an alarm is issued when oxygen levels are outside the acceptable range

1

Prototype

Quality Control OperationAIR

2020/04/07

passed

O2 alarm is issued as expected

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Connect device to test lung			
2	Connect 100% oxygen and air to device input			



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 46 / 75

3	Set minimum oxygen level to 60%			
4	Ventilate at PPLAT 35 cm H2O, PEEP 5 cm H2O, FiO2 80%			
5	Cut off oxygen input	Oxygen alarm is issued	passed	

### XTC-87 Air pressure drop alarm (TC-155)

TEST

Show that an alarm is issued when input air pressure is insufficient

1

Prototype

Quality Control OperationAIR

2020/04/07

passed

Oxygen too high alarm is issued

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Connect device to test lung			
2	Ventilate with PPLAT 25, PEEP 10, Freq 15, FiO2 40%			
3	Set maximum oxygen to 50%			
4	Disconnect input air			
5	wait 20 seconds	An alarm is issued	passed	

### XTC-88 Indication of parameters relevant to safety (TC-97)

Address the need for the indication of parameters that are associated with hazardous output in Risk Management Process (i.e. you want to know the air pressure, volume of air, oxygen content that is delivered to the patient)

1

Current RMF

Legal OperationAIR



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 47 / 75

2020/04/07				
passed				
Under folder H02				
Step	Action Result	Expected Result	Passed/Failed	Comment

### XTC-89 Usability (TC-94)

Risks of poor usability, including those associated with identification, marking and documents must be addressed in Usability engineering process results

1

Current RMC

Legal OperationAIR

2020/04/07

passed with deviation

This is elaborately tested over the entirety of the RMF. Still some changes can occur.

Step	Action Result	Expected Result	Passed/Failed	Comment

### XTC-90 Accuracy of controls and instruments (TC-93)

Risks associated with accuracy of controls and instruments must be addressed in Risk management file

1

Current RMF

Legal OperationAIR

2020/04/07



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 48 / 75

passed

Described in folder H03

Step	Action Result	Expected Result	Passed/Failed	Comment
------	---------------	-----------------	---------------	---------

### XTC-91 Push test (TC-61)

External parts of an ENCLOSURE are subject to a steady force of  $250\text{ N} \pm 10\text{ N}$  for a period of 5 s, applied by means of a suitable test tool providing contact over a circular plane surface 30 mm in diameter. However, this test is not applied to the bottom of an ENCLOSURE.

After the test, any damage sustained that results in an unacceptable RISK, as determined by inspection of the RISK MANAGEMENT FILE, constitutes a failure.

1

First prototype moulded encasing

Quality Control OperationAIR

2020/04/08

passed

Material is strong enough, no damage

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Apply force of $250\text{ N} \pm 10\text{ N}$ for a period of 5 s, over a circular plane surface 30 mm in diameter on the enclosure (not to the bottom)	No damage seen at inspection after push test	passed	

### XTC-92 Impact test (TC-62)

The medical device must withstand an impact of a solid smooth steel ball, approximately 50 mm in diameter and with a mass of  $500\text{ g} \pm 25\text{ g}$ , falling freely from 1,3 m height once onto each relevant part of the test sample or swinging like a pendulum, that drops through a vertical distance of 1,3 m, against vertical surfaces. The test is not applied to flat panel displays.

1



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 49 / 75

First prototype moulded encasing				
Quality Control OperationAIR				
2020/04/08				
passed				
No damage				
Step	Action Result	Expected Result	Passed/Failed	Comment
1	a solid smooth steel ball, approximately 50 mm in diameter and with a mass of 500 g ± 25 g, falling freely from 1,3 m height once onto each relevant part of the test sample or swinging like a pendulum, that drops through a vertical distance of 1,3 m, against vertical surfaces. The test is not applied to flat panel displays.	No damage is sustained that results in unacceptable risk	passed	

<b>XTC-93 Temperature of applied parts (TC-85)</b>				
Temperature of tube must be disclosed in Risk Management File when it exceeds 41 degrees Celsius or is below ambient temperature				
1				
Prototype				
Quality Control OperationAIR				
2020/04/08				
passed				
Temperature of air tube at patient side is 23.4 degrees celsius Ambient temperature 21.7 degrees celsius				
Step	Action Result	Expected Result	Passed/Failed	Comment



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 50 / 75

1	Connect device to test lung			
2	Ventilate with PPLAT 10, PEEP 5, Freq 10			
3	Measure temperature of intubation tube using infrared thermometer	temperature is ambient temperature + - 2 degrees celcius	passed	
4				

### XTC-94 Impairment of cooling (TC-99)

#### TEST

The medical device must remain safe during the failure of cooling systems, f.e. when ventilation openings are covered.

1

Prototype

Quality Control OperationAIR

2020/04/06

passed

Covered ventilation holes for 10 minutes.  
Temperatures in housing: ambient 30, Valve 52, GUI 51

Step	Action Result	Expected Result	Passed/Failed	Comment
1	connect device to test lung			
2	Ventilate with PPLAT 25, PEEP 10, Freq 30			
3	Cover ventilation holes and wait 10 minutes			
4	Measure temperature inside casing	does not exceed 60 degrees celcius	passed	

### XTC-96 Alarm settings (TC-157)

#### INSPECTION

Show that the alarm minimum and maximum settings can be set by the users

1



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 51 / 75

Prototype				
Quality Control OperationAIR				
2020/04/08				
passed				
Can be set by user				
Step	Action Result	Expected Result	Passed/Failed	Comment

**XTC-97 Alarms easy to understand (TC-119)**

INSPECT				
Making sure the alarms are made according to ISO 60601-1-8				
1				
Prototype				
Quality Control OperationAIR				
2020/04/08				
passed				
Passed				
Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect alarms with documentation	Alarms are according to 60601-1-8	passed	
2				
3				
4				

**XTC-98 Splash proof (TC-3)**

TEST				
Confirm that the housing is able to protect the internal components from ingress of water				



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 52 / 75

1				
ABS housing prototype				
Quality Control OperationAIR				
2020/04/09				
passed				
Passed				
Step	Action Result	Expected Result	Passed/Failed	Comment
1	Setup device with test lung and default ventilation settings			
2	Drip water with 3 mm/min on device placed at 15 degree	Device stays fully functional	passed	
3	Repeat for angles at 4 directions	Device stays fully functional	passed	
4	Check if any water has breached the external casing	No water is inside the casing	passed	

### XTC-99 Off switch protection (TC-21)

#### DEMONSTRATION

Show that the off switch is positioned so that it is protected from accidental switching

1				
Prototype				
Quality Control OperationAIR				
2020/04/09				
passed				
Off switch is located at power adapter box, relatively protected				



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 53 / 75

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Find location of off switch, try to bump into it		passed	

### XTC-100 Intentional exceeding of safety limits (TC-96)

Risks associated with hazardous output arising from the intentional exceeding of safety limits must be addressed in risk management file

1

Current RMF

Legal OperationAIR

2020/04/10

passed

Documented in Risk 75 H10.4

Step	Action Result	Expected Result	Passed/Failed	Comment
------	---------------	-----------------	---------------	---------

### XTC-101 PEEP alarm (TC-8)

TEST

Confirm that an alarm is issued when the PEEP decreases beneath the set value

1

Quality Control OperationAIR

2020/04/13

passed

React < 1 sec

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Connect device to MI lung simulator			



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 54 / 75

2	Set minimum PEEP to 15 cm H2O			
3	Ventilate with PPLAT 35 cm H2O, PEEP 10 cm H2O			
4	Wait 10 seconds	a PEEP alarm is issued	passed	

### XTC-102 Plateau pressure alarm II (TC-83)

TEST

Confirm that an alarm is issued when the plateau pressure is outside the acceptable range set by the user.

1

Quality Control OperationAIR

2020/04/13

passed

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Connect device to test lung			
2	Set maximum plateau pressure to 30 cm H2O			
3	Ventilate with PPLAT 35, PEEP 5 cm H2O, Freq 15/min, FiO2 21%	An alarm is issued indicating PPLAT is too high	passed	

### XTC-103 Apnea/Disconnect alarm (TC-101)

TEST

Confirm that an alarm is issued in case of apnea or disconnection of tubes

1

Quality Control OperationAIR

2020/04/13

passed

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Connect device to test lung			



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 55 / 75

2	Ventilate with PPLAT 25, PEEP 5 cm h2O, Freq 15/min, I:E 0.5			
3	Block expiration tube			
4	wait 10 seconds	an alarm is issued	passed	
5	Unblock expiration tube	Alarm stops	passed	
6	Block inspiration tube			
7	wait 10 seconds	An alarm is issued	passed	
8	disconnect inspiration tube			
9	wait 10 seconds	An alarm is issued	passed	
10	reconnect inspiration tube	alarm stops	passed	
11	disconnect expiration tube			
12	wait 10 seconds	An alarm is issued	passed	

### **XTC-104 Battery Housing (TC-122)**

#### INSPECTION

Address that battery housing has no ventilation and that the battery is medically certified in Risk Management File

1

Current RMF

Legal OperationAIR

2020/04/13

passed

Documented in Risk-76, H01.4

Step	Action Result	Expected Result	Passed/Failed	Comment
------	---------------	-----------------	---------------	---------

### **XTC-105 Arrangements of controls and indicators (TC-59)**

the MANUFACTURER shall address in the RISK MANAGEMENT PROCESS the RISKS associated with the arrangement of controls and indicators of ME EQUIPMENT.



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 56 / 75

1				
Current RMF				
Legal OperationAIR				
2020/04/13				
passed				
Is documented in the Risks under folder H02,H03, H06 and H09				
Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect risk management file	is present	passed	

### XTC-106 Fail-safe valve (TC-67)

TEST				
Confirm that the fail-safe valve opens and releases pressure when pressure exceeds 70 cm H2O				
2				
Prototype, fail-safe placed at the expiratory module side				
Quality Control OperationAIR				
2020/04/10				
passed				
Pressure peaks at 96 cm H2O, and then decreases and settles at 70 after 2 seconds				
Step	Action Result	Expected Result	Passed/Failed	Comment
1	Attach test lung to device			
2	Attach pressure sensor to the system at the side of the test lung			
3	open expiratory valve			



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 57 / 75

4	Open inspiratory valve to put 600 mbar on system			
5	Read out pressure sensor	Does not exceed 100 mbar on peak pressure, and settles at 70 mbar	passed	

### XTC-107 GUI is clearly visible with protective eyewear (TC-158)

To test whether the GUI is clearly visible, the subject has to change settings using protective eyewear, normally used on ICUs. These are usually made from PET.

1

AIRone Version1.1

QA OperationAIR

2020/04/17

passed

The GUI is clear to read with eyewear. There is no difficulty experienced in adjusting settings or interpreting measurements using eyewear or without.

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Put on protective eyewear		passed	
2	Turn on device		passed	
3	Adjust settings:		passed	
4	- PEEP		passed	
5	- Plateau Pressure		passed	
6	Interpret measurements	Everything is clear and easy to read	passed	

### XTC-108 INSPECT: Information about fire hazard on the label (TC-166)

INSPECT:  
Inspect the sticker

1

The first version of the sticker

Legal OperationAIR



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 58 / 75

2020/04/20

passed

Fire symbol is present

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect sticker for firesymbol	Fire is forbidden symbol is present	passed	

### XTC-109 Pressure setting (TC-168)

DEMONSTRATION

Show that the pressure above PEEP can be set by the user between 5 and 40 in intervals of 1 cm H2O

1

Prototype

Quality Control OperationAIR

2020/04/23

passed

Pressure setting as relative pressure above PEEP, can be set between 5 and 40, interval is 1

Step	Action Result	Expected Result	Passed/Failed	Comment
------	---------------	-----------------	---------------	---------

### XTC-110 Usability test (TC-65)

This Test Case is meant to test whether a potential user can easily comprehend the functionalities (GUI, inputs and outputs) of the device. The goal is to test if the device is intuitive and to see if untrained personnel can safely use the device.

This test uses the test protocol 'Protocol (pre-)klinische tests RMVS' developed by Peter Somhorst

1

Prototype



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 59 / 75

Quality Control OperationAIR

2020/04/10

**documentation**

See notes pre-clinical test 10-04-2020

Step	Action Result	Expected Result	Passed/Failed	Comment
1				

### **XTC-111 Expiratory Hold (TC-170)**

DEMONSTRATION

Show that the device can perform an expiratory hold

1

prototype

Quality Control OperationAIR

2020/04/24

passed

Step	Action Result	Expected Result	Passed/Failed	Comment

### **XTC-112 HEPA filter (TC-171)**

INSPECTION

Show that a HEPA filter can be equipped

1

Prototype

Quality Control OperationAIR

2020/04/24



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 60 / 75

passed

Expiratory module has a 22 mm connector which can be connected to a standard HEPA filter

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect design drawings			

### XTC-113 Clear explanation per alarm (TC-17)

DEMONSTRATION

When an alarm is issued, the user interface indicates what alarm is issued

1

Prototype

Quality Control OperationAIR

2020/04/24

passed

When an alarm is issued, the parameter outside of its limits is shown in red. Additionally, all previous alarms are displayed in the alarm overview until acknowledged by the user

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect user interface while device issues an alarm		passed	

### XTC-114 Airflow materials (TC-159)

INSPECTION

Show that all parts which have inspiratory air flowing through them are biocompatible and oxygen compatible

1

Documentation

Quality Control OperationAIR

2020/04/24



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 61 / 75

passed

Inspected Technical Manual. Most materials are proven biocompatible, and when not proven a rationale is provided

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect Technical Manual	Each part which has inspiratory air flowing through it has a description of biocompatibility and oxygen compatibility	passed	

### XTC-115 Cleaning and disinfection (TC-127)

TEST

Confirm that the device can withstand regular cleaning and disinfection with alcohol 70%

1

Final design casing and labeling

Quality Control OperationAIR

2020/04/24

passed

No deterioration seen in casing material or labeling

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Subject device to 14 cleaning cycles with alcohol 70%, wait until alcohol dries inbetween cycles. Make sure to clean the labels and touchscreen as well	Device is not affected by cleaning cycles. Labels do not degrade	passed	

### XTC-116 Pressure Hold (TC-91)

TEST

Confirm that the device does not leak oxygen by keeping the airflow system pressurized for an extended period of time



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 62 / 75

2

Prototype

Quality Control OperationAIR

2020/04/24

**failed**

Pressure loss too large

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Connect device to test lung			
2	Block the expiration tube			
3	Pressurize the tubes to 70 cm H2O			
4	Hold pressure for 2 minutes and continuously read out the internal pressure sensor			
5	Analyse pressure loss over time	Pressure loss in 2 minutes is smaller than 10 cm H2O	<b>failed</b>	
6				

### XTC-117 Sharp edges (TC-48)

TEST

Confirm that the device has no sharp edges

1

Final housing

Quality Control OperationAIR

2020/04/24

passed

Passed



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 63 / 75

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Put on protective glove on one hand			
2	Trace hand with glove along all edges of the device	Glove does not tear	passed	
3				

### XTC-118 Tipping test II (TC-110)

TEST

Show that the device is able to withstand unintended forces without tipping over.

1

Prototype

Quality Control OperationAIR

2020/04/24

passed with deviation

Does not fall over. With force applied from some directions, the device does displace more than 50 mm, because it only has brakes on the two front wheels

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Put device on horizontal plane			
2	Put device on brakes			
3	Apply a force of 25% of the weight of the device parallel to the horizontal at 1.5m height	Device does not fall over; Does not displace more than 50 mm	passed	

### XTC-119 Usability with protective equipment (TC-41)

TEST

Confirm that the user interface touchscreen works when the user wears protective gloves and can easily be seen when wearing protective eyewear.

2

Prototype

Quality Control OperationAIR



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 64 / 75

2020/04/24

passed

Touchscreen works also when wearing gloves, and can be seen while wearing protective eyewear

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Put on two layers of protective gloves			
2	Put on protective glasses			
3	Adjust several settings on the user interface	Touchscreen responds to touching	passed	

### XTC-120 Internal components tightly attached (TC-4)

TEST

Confirm the device can withstand basic movements expected during transport and all components stay attached

1

Prototype

Quality Control OperationAIR

2020/04/24

passed with deviation

Removed device from frame to make shaking easier

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Shake device manually for 1 minute			
2	Remove housing			
3	Look & Feel all internal components	Everything is attached/ doesn't move	passed	

### XTC-121 Visibility of label (TC-160)

DEMONSTRATION

Show that the label can be seen and read while standing next to the device at 1 m distance



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 65 / 75

1					
Final housing and labeling					
Quality Control OperationAIR					
2020/04/24					
passed					
Label is placed at the back of the device. Can be seen when device is turned around					
<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th style="width: 10%;">Step</th> <th style="width: 25%;">Action Result</th> <th style="width: 25%;">Expected Result</th> <th style="width: 15%;">Passed/Failed</th> <th style="width: 25%;">Comment</th> </tr> </thead> </table>	Step	Action Result	Expected Result	Passed/Failed	Comment
Step	Action Result	Expected Result	Passed/Failed	Comment	

### XTC-132 Installation Verification (TC-174)

INSPECTION				
Inspect whether the necessary tests are incorporated in the Installation-Verification protocol of the AIRone.				
1				
Approved 1 of the IV-INT.AIRone				
Legal OperationAIR				
2020/04/24				
passed				
All tests are present.				
Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect the Installation-Verification protocol if the following tests are present:			
2	Serial numbers of components have to be registered	Is present	passed	IV-INT.AIRone4
3	The power supply has to be inspected	Is present	passed	IV-INT.AIRone5
4	The device has to be inspected visually	Is present	passed	IV-INT.AIRone6



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 66 / 75

5	Pressure sensors have to be inspected	Is present	passed	IV-INT.AIRone7
6	The device has to be turned on and off and power down has to be inspected	Is present	passed	IV-INT.AIRone8
7	Electrical safety has to be inspected	Is present	passed	IV-INT.AIRone9
8	Sensors, valves and output have to be tested	Is present	passed	IV-INT.AIRone10
9	Gas system integrity has to be tested.	Is present	passed	IV-INT.AIRone11

### XTC-122 Electromagnetic compatibility (TC-58)

Risks associates with

- electromagnetic phenomena existing at the locations where the ME EQUIPMENT or ME SYSTEM is intended to be used
- the introduction by the ME EQUIPMENT or ME SYSTEM of electromagnetic phenomena into the environment that might degrade the performance of other devices, electrical equipment and systems must be addressed in Risk Management Process

1

Documentation

Quality Control OperationAIR

2020/04/24

passed

Documented in risks 80 and 81

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect Risk Management file		passed	

### XTC-123 Ventilator (TC-172)

INSPECTION

Show that the device has a ventilator that can displace potential leaking oxygen to outside the encasing

1

Prototype



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 67 / 75

Quality Control OperationAIR				
2020/04/24				
passed				
Device has a ventilator in the bottomplate				
Step	Action Result	Expected Result	Passed/Failed	Comment

<b>XTC-124 Tidal volume alarm I (TC-7)</b>				
Confirm that an alarm is issued when tidal volume is outside acceptable range				
1				
Prototype				
Quality Control OperationAIR				
2020/04/24				
passed				
Tidal volume too low alarm is issued				
Step	Action Result	Expected Result	Passed/Failed	Comment
1	Connect device to MI lung simulator			
2	Set minimum tidal volume to 300 mL			
3	ventilate with Ppeak 15 cm H <sub>2</sub> O, PEEP 5 cm H <sub>2</sub> O			
4	Wait 10 seconds	A tidal volume alarm is provided	passed	

<b>XTC-125 Tidal volume alarm II (TC-68)</b>				
Confirm that an alarm is given when tidal volume is outside acceptable range				
1				



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 68 / 75

Prototype				
Quality Control OperationAIR				
2020/04/24				
passed				
Tidal volume too high alarm is issued				
Step	Action Result	Expected Result	Passed/Failed	Comment
1	Connect device to MI lung simulator			
2	Set lung compliance to 0.03			
3	Set maximum tidal volume to 300 mL			
4	ventilate with Ppeak 30 cm H2O, PEEP 5 cm H2O			
5	Wait 10 seconds	A tidal volume alarm is provided	passed	

### XTC-126 HEPA filter warning sticker (TC-173)

#### INSPECTION

Show that the device has a warning sticker above the expiratory air connector to not forget to connect a HEPA filter

1

Final housing and labeling

Quality Control OperationAIR

2020/04/24

passed

Passed

Step	Action Result	Expected Result	Passed/Failed	Comment
------	---------------	-----------------	---------------	---------



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 69 / 75

### XTC-127 IFU packed with device (TC-162)

**DEMONSTRATION**

Show that the IFU is available for download online, and show packaging list includes a printed IFU

1

Documentation

Quality Control OperationAIR

2020/04/24

passed

Step	Action Result	Expected Result	Passed/Failed	Comment
------	---------------	-----------------	---------------	---------

### XTC-128 Back-up battery empty alarm (TC-84)

**TEST**

Confirm that an alarm is issued when the back-up battery is almost empty

1

Prototype

Quality Control OperationAIR

2020/04/24

**failed**

Alarm is directly issued when mains power is disconnected. Battery percentage shown is 0%, while battery is fully charged

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Connect device to test lung			
2	Disconnect power supply mains			
3	Ventilate with Ppeak 60, PEEP 5 cm H2O, Freq 30/min, FiO2 21%			



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 70 / 75

4	Keep ventilating until battery almost runs out	An alarm is issued	failed	
---	--	--------------------	--------	--

### XTC-129 Back up battery warning (TC-69)

Confirm that the device indicates when it is running on the back-up battery power source

1

Prototype

Quality Control OperationAIR

2020/04/24

failed

Device immediately alarms for empty battery, while battery is fully charged

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Connect device to test lung			
2	Ventilate with Ppeak 20 cm H2O, PEEP 10 cm H2O			
3	Cut off mains power supply	Device displays a warning that it is running on battery power supply	failed	

### XTC-130 Measurement Reliability (TC-106)

TEST

Confirm that the same ventilation settings at different times lead to the same output

1

Prototype

Quality Control OperationAIR

2020/04/24

passed



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 71 / 75

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Alternate the following settings: ventilate with Ppeak 25 cm H2O, PEEP 5 cm H2O, Freq 20, I:E 0.5, FiO2 21% at least 10 times with different settings throughout the testing period.	The set variables result in values that change maximally 5%	passed	

### XTC-131 Flow Sensor (TC-169)

#### DEMONSTRATION

Show that all conductive components near the flow sensor are grounded to the bottomplate

1

Documentation

Quality Control OperationAIR

2020/04/24

passed

Most components near the flowsensor are non-conductive. The conductive components are grounded

Step	Action Result	Expected Result	Passed/Failed	Comment
------	---------------	-----------------	---------------	---------

### XTC-133 Open Source (TC-175)

Documentations are Open source available via <https://osf.io/mn7xq/>

1

Legal OperationAIR

2020/04/24

passed

Documentations are Open source available via <https://osf.io/mn7xq/>

Step	Action Result	Expected Result	Passed/Failed	Comment
------	---------------	-----------------	---------------	---------



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 72 / 75

### XTC-134 Fail safe valve test during production (TC-111)

**INSPECTION**

Show that the Production Documentation mention testing each fail safe valve

1

Documentation

Quality Control OperationAIR

2020/04/28

passed

Stated in IV.Z8113020xx A4

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect production documentation		passed	

### XTC-135 IFU: regulatory requirements (TC-165)

The instructions must contain:

- the name of the device;
- the manufacturer of the device;
- the adress of the manufacturer;
- Indications and contra-indications,
- intended use and intended user;
- the clinical benefit for the patient;
- the performance characteristics;
- the degree of accuracy for the parameters of the device;
- how to verify if the right accessories are used;
- undesirable side-effects;
- preperation of the device before use;
- that users must receive a training before using the device;
- information on maintenance and cleaning of the device and accessories;
- information on cleaning and disinfection between the use on different patients;
- all the warnings that will lead to malfunction of the device and/or changes in the working of the device that will lead to risks to the patient. This will include enviromental changes, like diagnostic or therapeutic procedures;



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 73 / 75

- how the device can safely be disposed;
- publication or printing date with version number;
- that SAE's must be mentioned to the manufacturer;
- how to safely use the software on the device.

1

IFU v1.5

Quality Control OperationAIR

2020/04/29

passed with deviation

All requirements are filled. SAE's are not mentioned as such, but the instruction to contact the manufacturer in case of problems is.

Step	Action Result	Expected Result	Passed/Failed	Comment
------	---------------	-----------------	---------------	---------

### XTC-136 IFU Review (TC-43)

DEMONSTRATION

Show that the IFU has been checked by experts

1

IFU v 1.5

Quality Control OperationAIR

2020/04/29

passed

Usability approval  
QC approval



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 74 / 75

Step	Action Result	Expected Result	Passed/Failed	Comment
1				
2				
3				

### XTC-137 Label: regulatory requirements (TC-164)

The label must contain:

- the device name;
- a serial number;
- the manufacturer with address;
- the manufacturing date, as part of the serial number or separately;
- immediate warnings with icons.

All will be written in text or official icons.

1

Label

Quality Control OperationAIR

2020/04/28

passed

Passed

Step	Action Result	Expected Result	Passed/Failed	Comment
------	---------------	-----------------	---------------	---------

### XTC-138 Vocabulary and Semantics (TC-71)

INSPECTION

Confirm that the nomenclature, definitions, vocabulary and semantics used in all user communication correspond with BS ISO 19223:2019

1

Documentation IFU 1.5, final label

Quality Control OperationAIR



## Design Verification

OPERATIONAIR: OperationAIR  
(OperationAIR)

Date: 2020/04/29 14:28:42

Document: 4.1

Page: 75 / 75

2020/04/29

passed

Passed

Step	Action Result	Expected Result	Passed/Failed	Comment
1	Inspect IFU, GUI and training materials		passed	